

APPLICANT

Opharm Sp. z o.o.
POKRZYWNICA 62
99-120 Piatek .

SAMPLE DESCRIPTION Quotation No. 1451/IR
Sample Received 28/10/2020
Reference: K-1
Colour: WHITE
Batch number: K-11/2020
Fiber composition POLYPROPYLENE
Buyer's Name: Opharm Sp. z o.o.

Sample 01

Test	Rating	Failure result
Bacterial Filtration Efficiency EN 14683: 2019 + AC 2019	P	
Differential Pressure (BREATHABILITY) EN 14683:2019 + AC 2019	P	
Splash Resistance Pressure ISO 22609:2004	P	
Microbial Cleanliness/Bioburden EN 14683:2019/ EN ISO 11737-1:2018	P	

P = MEETS REQUIREMENT / A = ACCEPTABLE / F = DOES NOT MEET REQUIREMENT / NR = NO REQUIREMENT / SC = STILL CONTINUES / X = NOT PERFORMED
Values in brackets represent the limits



TEST METHOD	SAMPLE DESCRIPTION	PARAMETERS	RESULT	U.M.	LIMITS	NOTES	LAB
Bacterial Filtration Efficiency EN 14683: 2019 + AC 2019 <u>Operating Conditions</u> Test Conditions: Temperature: 21±5°C Humidity: 85±5% Dimensions of the test specimens: 49cm ² (5 test specimens) Side of the test specimen facing the challenge aerosol: intern Air flow rate: 28.3 l/min. MPS 2.9 The expanded uncertainty at a confidence level of 95%, k=2: 1.8%	01		Specimen 1 Specimen 2 Specimen 3 Specimen 4 Specimen 5 Result	98,7 98,7 98,8 99,0 98,5 98,7	% % % % % %	>=95	PT PT PT PT PT PT
Differential Pressure (BREATHABILITY) EN 14683:2019 + AC 2019 <u>Operating Conditions</u> Test Conditions: Temperature: 21±5°C Humidity: 85±5% Number and general location of the areas of the mask the differential measurements were taken: Test performed with the direction of flow from the inside to the outside. Side and central location. Air flow rate: 8L/min Dimensions of the test specimens: 4.9cm ² (5 test specimens) The expanded uncertainty at a confidence level of 95%, k=2: 8.7%	01		Specimen 1 Specimen 2 Specimen 3 Specimen 4 Specimen 5 Result	26,5 26,5 26,5 26,5 26,5 26,5	Pa/cm ² Pa/cm ² Pa/cm ² Pa/cm ² Pa/cm ² Pa/cm ²	<40	PT PT PT PT PT PT

Laboratory Manager


 Ilenia Francioni
 ilenia.francioni@intertek.com

Validated by: Ilenia Francioni

TEST METHOD	SAMPLE DESCRIPTION	PARAMETERS	RESULT	U.M.	LIMITS	NOTES	LAB
Splash Resistance Pressure ISO 22609:2004 <u>Operating Conditions</u> Test Conditions: Samples exposed to a jet of 2mL synthetic blood at pressure (low: 10.6 KPa; medium: 16.0 KPa; high:21.3 KPa) aimed at the centre of the mask. Test performed at laboratory temperature of 21°C and 45% relative humidity, within 60 seconds after the mask was removed from the conditioning chamber Observation after 10+1 second of blood penetration on the opposite side of the mask. Synthetic blood according to Annex B of ISO 22609: 2004 with surface tension of 42 + 2mN / Number and General location of the areas: 32 test specimen / center (pass at least Medium pressure test for 29 out of 32 samples as minimum, corresponding to AQL 4%, according EN 14683: 2019 mask Type IIR)	01		Result	16	kPa	>=16	PT
Microbial Cleanliness/Bioburden EN 14683:2019/ EN ISO 11737-1:2018 <u>Operating Conditions</u> Test Conditions: 5 min shaker at 250rpm Area of each test specimen: 5 test specimens Mic30°C (3 days), Molds and yeasts 25°C (7 days) The expanded uncertainty at a confidence level of 95%, k=2: 20%	01		Specimen 1 Specimen 2 Specimen 3 Specimen 4 Specimen 5 Result	19 23 24 20 19 21	UFC/g UFC/g UFC/g UFC/g UFC/g UFC/g	<=30	PT PT PT PT PT PT

Test performed by:

PT: Intertek Portugal Unipessoal LDA
 Rua Antero De Quental 221 S.102
 Perafita-Matosinhos

Amendment 1: splash test requirement was added as per applicant request
 This Report replaced the report n° ITAT20016345 dated on 13/11/2020 and must be used instead of it.

Laboratory Manager


 Ilenia Francioni
 ilenia.francioni@intertek.com

Validated by: Ilenia Francioni

The partial reproduction of the present Test Report is not allowed without written authorization of the General Manager.
The content of the present document refers exclusively to the submitted samples and not to the batch that they want to represent.
A representative portion of the sample is kept for 60 days. Specimen are kept for 30 days.
Samples are drawn by customer.
UM: unit of measurement
Limits are based on Customer's Manual in force available at Intertek Italia S.p.a. - Via di Stagno 17 F/G - 50055 Lastra a Signa (FI)
Uncertainty is the extended uncertainty calculated using a coverage factor of 2 which gives a level of confidence of approximately 95%.
All the results are reported without considering the uncertainty.
LQ: Quantification Limit
As received by the customer, the sample was accepted for testing.

RESULTS MARKED IN RED EXCEED LIMITS

END OF TEST REPORT

Laboratory Manager


Ilenia Francioni
ilenia.francioni@intertek.com

Validated by: Ilenia Francioni